

Chulalongkorn University

## Chula Digital Collections

---

Chulalongkorn University Theses and Dissertations (Chula ETD)

---

2020

# The effectiveness of led toothbrush in reducing dental plaque and gingivitis in fixed orthodontic patients: a randomized clinical trials

Chavirakarn Manphibool  
*Faculty of Dentistry*

Follow this and additional works at: <https://digital.car.chula.ac.th/chulaetd>



Part of the [Orthodontics and Orthodontology Commons](#)

---

### Recommended Citation

Manphibool, Chavirakarn, "The effectiveness of led toothbrush in reducing dental plaque and gingivitis in fixed orthodontic patients: a randomized clinical trials" (2020). *Chulalongkorn University Theses and Dissertations (Chula ETD)*. 62.

<https://digital.car.chula.ac.th/chulaetd/62>

This Thesis is brought to you for free and open access by Chula Digital Collections. It has been accepted for inclusion in Chulalongkorn University Theses and Dissertations (Chula ETD) by an authorized administrator of Chula Digital Collections. For more information, please contact [ChulaDC@car.chula.ac.th](mailto:ChulaDC@car.chula.ac.th).

THE EFFECTIVENESS OF LED TOOTHBRUSH IN REDUCING DENTAL PLAQUE AND  
GINGIVITIS IN FIXED ORTHODONTIC PATIENTS: A RANDOMIZED CLINICAL TRIALS



Miss Chavirakarn Manphibool

A Thesis Submitted in Partial Fulfillment of the Requirements  
for the Degree of Master of Science in Orthodontics

Department of Orthodontics

FACULTY OF DENTISTRY

Chulalongkorn University

Academic Year 2020

Copyright of Chulalongkorn University

ประสิทธิภาพของแปรงสีฟันแอลอีดีในการลดคราบจุลินทรีย์และภาวะเหงือกอักเสบในผู้ป่วยที่  
รักษาด้วยเครื่องมือจัดฟันชนิดติดแน่น



น.ส.ชวริกาญ์ แม่นพิบูลย์

วิทยานิพนธ์นี้เป็นส่วนหนึ่งของการศึกษาตามหลักสูตรปริญญาวิทยาศาสตรมหาบัณฑิต

สาขาวิชาทันตกรรมจัดฟัน ภาควิชาทันตกรรมจัดฟัน

คณะทันตแพทยศาสตร์ จุฬาลงกรณ์มหาวิทยาลัย

ปีการศึกษา 2563

ลิขสิทธิ์ของจุฬาลงกรณ์มหาวิทยาลัย

Thesis Title	THE EFFECTIVENESS OF LED TOOTHBRUSH IN REDUCING DENTAL PLAQUE AND GINGIVITIS IN FIXED ORTHODONTIC PATIENTS: A RANDOMIZED CLINICAL TRIALS
By	Miss Chavirakarn Manphibool
Field of Study	Orthodontics
Thesis Advisor	Assistant Professor PINTUON CHANTARAWARATIT, D.D.S., Ph.D.
Thesis Co Advisor	Associate Professor ORANART MATANGKASOMBUT, D.D.S., Ph.D.

---

Accepted by the FACULTY OF DENTISTRY, Chulalongkorn University in Partial  
Fulfillment of the Requirement for the Master of Science

..... Dean of the FACULTY OF DENTISTRY  
(Associate Professor Pornchai Jansisyanont, D.D.S., M.S.,  
Ph.D.)

#### THESIS COMMITTEE

..... Chairman  
(SIRICHOM SATRAWAHA, D.D.S., Ph.D.)

..... Thesis Advisor  
(Assistant Professor PINTUON CHANTARAWARATIT, D.D.S.,  
Ph.D.)

..... Thesis Co-Advisor  
(Associate Professor ORANART MATANGKASOMBUT, D.D.S.,  
Ph.D.)

..... External Examiner  
(Associate Professor Poonsak Pisek, D.D.S., Ph.D.)

ชวிரากญ แม่นพิบูลย์ : ประสิทธิภาพของแปรงสีฟันแอลอีดีในการลดคราบจุลินทรีย์และภาวะ  
เหงือกอักเสบในผู้ป่วยที่รักษาด้วยเครื่องมือจัดฟันชนิดติดแน่น. ( THE EFFECTIVENESS OF  
LED TOOTHBRUSH IN REDUCING DENTAL PLAQUE AND GINGIVITIS IN FIXED  
ORTHODONTIC PATIENTS: A RANDOMIZED CLINICAL TRIALS) อ.ที่ปรึกษาหลัก : ผศ.  
ทญ. ดร.พินทุอร จันทรวราทิพย์, อ.ที่ปรึกษาร่วม : รศ. ทญ. ดร.อรนาฏ มาตั้งคสมบัติ

วัตถุประสงค์: เพื่อเปรียบเทียบประสิทธิภาพของแปรงสีฟันแอลอีดีและแปรงสีฟันธรรมดาต่อการ  
ลดคราบจุลินทรีย์และภาวะเหงือกอักเสบในผู้ป่วยที่จัดฟันด้วยเครื่องมือจัดฟันชนิดติดแน่น และเพื่อทดสอบ  
ประสิทธิภาพของแปรงสีฟันแอลอีดีรวมถึงผลของระยะเวลาได้รับแสงที่แตกต่างกันต่อแผ่นชีวภาพของ  
แบคทีเรียสเตรปโตคอคคัสมิวแทนส์ในห้องปฏิบัติการ

วัสดุและวิธีการ: ผู้ป่วยที่จัดฟันด้วยเครื่องมือจัดฟันชนิดติดแน่น 15 รายได้รับการจัดกลุ่มแบบสุ่ม  
โดยแบ่งเป็นกลุ่มที่ได้รับแปรงสีฟันแอลอีดี และกลุ่มที่ได้รับแปรงสีฟันธรรมดา โดยจะได้รับการวัดค่าดัชนี  
คราบจุลินทรีย์และดัชนีสภาพเหงือกก่อนได้รับแปรงและหลังจากใช้แปรงไปแล้ว 28 วัน การทดลองใน  
ห้องปฏิบัติการแผ่นชีวภาพของแบคทีเรียสเตรปโตคอคคัสมิวแทนส์จะถูกแบ่งเป็น 5 กลุ่ม กลุ่มละ 6 ตัวอย่าง  
ตามระยะเวลาที่ได้รับแสงแอลอีดีได้แก่ 15 วินาที, 30 วินาที, 60 วินาที, 120 วินาที และกลุ่มที่ไม่ได้รับแสง  
แอลอีดีเลย

ผลการทดลอง: ค่าดัชนีคราบจุลินทรีย์และค่าดัชนีสภาพเหงือกระหว่างกลุ่มที่ใช้แปรงสีฟันแอลอีดี  
และกลุ่มที่ใช้แปรงสีฟันธรรมดาไม่แตกต่างกันอย่างมีนัยสำคัญ จากการทดลองในห้องปฏิบัติการพบว่ากลุ่ม  
แผ่นชีวภาพที่ได้รับแสงแอลอีดีเป็นระยะเวลา 15 วินาที, 30 วินาที, 60 วินาที และ 120 วินาที มีการลดลงของ  
ความมีชีวิตของเชื้อสเตรปโตคอคคัสมิวแทนส์มากกว่ากลุ่มที่ไม่ได้รับแสงแอลอีดีอย่างมีนัยสำคัญ

สรุปผลการทดลอง: แปรงสีฟันแอลอีดีไม่มีประสิทธิภาพในการลดคราบจุลินทรีย์และภาวะเหงือก  
อักเสบในผู้ป่วยที่จัดฟันด้วยเครื่องมือจัดฟันชนิดติดแน่นมากกว่าแปรงสีฟันธรรมดา การทดสอบใน  
ห้องปฏิบัติการพบว่าแสงสีฟ้าจากแปรงแอลอีดีสามารถลดความมีชีวิตของเชื้อสเตรปโตคอคคัสมิวแทนส์ใน  
แผ่นชีวภาพได้เมื่อได้รับแสงแอลอีดีเป็นระยะเวลาอย่างน้อย 15 วินาที

สาขาวิชา      ทันตกรรมจัดฟัน  
ปีการศึกษา    2563

ลายมือชื่อนิสิต .....  
ลายมือชื่อ อ.ที่ปรึกษาหลัก .....  
ลายมือชื่อ อ.ที่ปรึกษาร่วม .....

# # 6270024932 : MAJOR ORTHODONTICS

KEYWORD: Dental plaque removal, Fixed orthodontic patients, Blue light, Biofilm, LED toothbrush, Streptococcus mutans

Chavirakarn Manphibool : THE EFFECTIVENESS OF LED TOOTHBRUSH IN REDUCING DENTAL PLAQUE AND GINGIVITIS IN FIXED ORTHODONTIC PATIENTS: A RANDOMIZED CLINICAL TRIALS. Advisor: Asst. Prof. PINTUON CHANTARAWARATIT, D.D.S., Ph.D. Co-advisor: Assoc. Prof. ORANART MATANGKASOMBUT, D.D.S., Ph.D.

Objective: To compare the effectiveness of LED toothbrush to manual toothbrush in reducing dental plaque and gingival inflammation in fixed orthodontic patients, and to investigate the effect of duration of LED toothbrush exposed to the *S.mutans* biofilm *in vitro*.

Materials and methods: Fifteen fixed orthodontic patients were recruited to this parallel-group analysis. The patients were randomly divided into 2 groups relying on brushing methods: manual toothbrush and LED toothbrush. Plaque index and gingival index were examined by a calibrated-blinded examiner at baseline and 28 days after brushing period. *In vitro* part, the *S. mutans* biofilms were assigned to 5 groups with 6 samples each, depending on the duration of LED exposure, which are 15 seconds, 30 seconds, 60 seconds, 120 seconds, and the control with no LED exposure.

Results: Between-group comparisons showed no significant difference in plaque index and gingival index. The LED toothbrush significantly reduced dental plaque at the gingival portion on the bracket side. *In vitro* part, the percentage of bacterial viability was significantly reduced in 15, 30, 60, 120 seconds group.

Conclusion: LED toothbrush did not more effective in reducing dental plaque and gingival inflammation than the manual toothbrush in fixed orthodontic patients. The LED blue light from the LED toothbrush significantly reduced the number of *S.mutans* in biofilm *in vitro* when the biofilm was exposed to the light for at least 15 seconds.

Field of Study: Orthodontics

Student's Signature .....

Academic Year: 2020

Advisor's Signature .....

Co-advisor's Signature .....

## ACKNOWLEDGEMENTS

I would like to take this opportunity to express my gratitude to Assist. Prof. Dr. Pintu-on Chantarawaratit and Assoc. Prof. Dr. Oranart Matangkasombut for their precious time and dedicated guidance. I also appreciate the kind suggestion from colleagues at the Department of Orthodontics and Department of Microbiology, Faculty of Dentistry, Chulalongkorn University. Moreover, I am grateful to Assoc. Prof. Dr. Pichayada Ketmake, Department of Imaging and Printing Technology, Faculty of Science, Chulalongkorn University and Assoc. Prof. Dr. Thavatchai Tayjasanant, Department of Electrical Engineering, Faculty of Engineer, Chulalongkorn University for their valuable supports for the radiance measurement of the LED toothbrush. This work is reinforced by the Ratchadapiseksomphot endowment fund of Chulalongkorn University (to Research Unit on Oral Microbiology and Immunology).

Chavirakarn Manphibool



## TABLE OF CONTENTS

	Page
.....	iii
ABSTRACT (THAI) .....	iii
.....	iv
ABSTRACT (ENGLISH) .....	iv
ACKNOWLEDGEMENTS.....	v
TABLE OF CONTENTS.....	vi
LIST OF TABLES.....	ix
LIST OF FIGURES .....	x
CHAPTER 1 INTRODUCTION .....	1
Background and Rationale .....	1
Research Questions.....	2
Research Objectives.....	2
Research Hypothesis.....	2
Null hypothesis: .....	2
Alternative hypothesis: .....	3
<u>Significance of the Study</u> .....	3
Conceptual Framework.....	4
CHAPTER 2 LITERATURE REVIEW.....	5
Dental Plaque .....	5
Oral microorganisms in orthodontic patients .....	7
Plaque removal in orthodontic patients .....	8



Photodynamic therapy, Light-emitting diode (LED) - Blue light and antimicrobial activity .....	9
Light-emitting diode (LED) toothbrushes.....	10
CHAPTER 3 MATERIALS AND METHODS .....	12
Participants .....	12
Sample size .....	12
Allocation technique .....	13
Ethical consideration.....	14
LED toothbrush .....	14
Research methodology.....	14
1. The clinical part .....	14
2. The <i>in vitro</i> part.....	15
Outcome measurement .....	17
Data collection and analysis .....	19
Expected benefit of this study .....	20
Limitation .....	20
CHAPTER 4 RESULT .....	20
The clinical part.....	20
Plaque index score.....	20
Gingival index score.....	22
The <i>in vitro</i> part .....	23
CHAPTER 5 DISCUSSION .....	25
APPENDICE .....	28


Appendix A .....	28
A.1 The individual mean Plaque Index Score (PI) of bracket and non-bracket sides before and after brushing by each method. ....	28
A.2 The individual mean Plaque Index Score (PI) of a different tooth surface area around the bracket on the bracket side before and after brushing by each method. ....	29
A.3 The individual mean Gingival Index Score (GI) of bracket and non-bracket sides before and after brushing by each method. ....	30
A.4 The geometric mean of the numbers of <i>S.mutans</i> and the calculation to percentage of bacterial viability. ....	31
Appendix B .....	33
REFERENCES.....	35
VITA .....	43

## LIST OF TABLES

	Page
Table 1 Group allocation of 25 subjects .....	13
Table 2 The Plaque index system .....	18
Table 3 Criteria for the Gingival Index Score .....	19
Table 4 Mean (SD) Plaque Index Score (PI) of bracket and non-bracket sides before and after brushing by each method and their differences. ....	21
Table 5 Mean (SD) Plaque Index Score (PI) of a different tooth surface area around the bracket on the bracket side before and after brushing by each method and their differences. ....	22
Table 6 Mean (SD) Gingival Index Score (GI) of bracket and non-bracket side before and after brushing by each method and their differences. ....	23
Table 7 Percentage of bacterial viability <sup>a</sup> .....	24

## LIST OF FIGURES

	Page
Figure 1 The conceptual framework .....	4
Figure 2 Table of computer-generated random numbers .....	13
Figure 3 The cross-over study design flow chart.....	15
Figure 4 Flow chart of the in vitro study design .....	17
Figure 5 Plaque index bracket.....	18
Figure 6 The comparison of percentage of bacterial viability .....	24



## CHAPTER 1 INTRODUCTION

### Background and Rationale

Dental plaque is the main etiology of dental caries and periodontal disease(1), which are challenging to a long time procedure like orthodontic treatment. Fixed orthodontic appliances have a complex shape of bracket wing and slot that are likely to be the plaque retentive area. Evidence showed that patients with fixed orthodontic appliances have higher plaque accumulation(2-6), the higher tendency of gingival inflammation, and bleeding on probing values(2, 4, 5, 7-9). Also, fixed orthodontic appliances cause the microbial shift towards more pathogenic types, for example, streptococci and lactobacilli(10) which are cariogenic bacteria, and the periodontopathogens like *T.forsythia*, *C.rectus*, and *P.nigrescens*(11). Hence, plaque control in fixed orthodontic patients is undeniably important. Due to the difficulties in plaque removal, a power toothbrush is one of the attempts trying to improve oral hygiene in fixed orthodontic patients. However, the evidence of its effectiveness is still in controversy(12, 13)

Recently, a light emitting diode (LED) toothbrush is trying to apply the photodynamic effect on the dental plaque in addition to mechanical effect. Because blue light was proved to have the bactericidal effect on *S. mutans* and oral biofilms(14-17), it is possible that an LED toothbrush would be more effective than a manual toothbrush, supported by a few studies. Clinically, the blue LED toothbrushes with 412 nm wavelength significantly reduced dental plaque, gingival bleeding, and inflammation more than the manual toothbrushes(18). However, the study in this area is still limited, and yet in orthodontic patients. The aim of this study is to investigate the effectiveness of LED toothbrush in dental plaque removal, reduction of gingival inflammation, and the number of mutans streptococci in fixed orthodontic patients.

### Research Questions

1. Is an LED toothbrush more effective than a manual toothbrush in removing dental plaque in fixed orthodontic patients?
2. Is an LED toothbrush more effective than a manual toothbrush in reducing gingival inflammation in fixed orthodontic patients?
3. Does an LED toothbrush have a bactericidal effect on newly formed *Streptococcus mutans* biofilm *in vitro*?
4. Do the different exposure times of an LED toothbrush have the different bactericidal effect on *Streptococcus mutans* biofilm *in vitro*?

### Research Objectives

1. To compare the effectiveness of LED toothbrushes and manual toothbrushes in dental plaque removal in fixed orthodontic patients.
2. To compare the effectiveness of LED toothbrushes and manual toothbrushes in the reduction of gingival inflammation in fixed orthodontic patients.
3. To evaluate whether an LED toothbrush have a bactericidal effect on newly formed *Streptococcus mutans* biofilm *in vitro*.
4. To evaluate whether the different exposure times of an LED toothbrush have the different bactericidal effect on *Streptococcus mutans* biofilm *in vitro*?

### Research Hypothesis

Null hypothesis:

1. An LED toothbrush is not more effective than the manual toothbrush in dental plaque removal in fixed orthodontic patients.
2. An LED toothbrush is not more effective than the manual toothbrush in the reduction of gingival inflammation in fixed orthodontic patients.

3. An LED toothbrush does not have a bactericidal effect on newly formed *Streptococcus mutans* biofilm *in vitro*.
4. The different exposure times of an LED toothbrush have no different bactericidal effect on *Streptococcus mutans* biofilm *in vitro*.

Alternative hypothesis:

1. An LED toothbrush is more effective than the manual toothbrush in dental plaque removal in fixed orthodontic patients.
2. An LED toothbrush is more effective than the manual toothbrush in the reduction of gingival inflammation in fixed orthodontic patients.
3. An LED toothbrush has a bactericidal effect on newly formed *Streptococcus mutans* biofilm *in vitro*.
4. The different exposure times of an LED toothbrush have the different bactericidal effect on *Streptococcus mutans* biofilm *in vitro*.

#### Significance of the Study

The result of a randomized controlled trial with a parallel-group design which investigate whether the LED electric toothbrush is beneficial in reducing dental plaque and improving gingival condition in patients with fixed orthodontic appliances.

## Conceptual Framework




Figure 1 The conceptual framework

The conceptual framework shows independent variable (LED toothbrush), dependent variables (Reduction in plaque, gingival inflammation, and mutans streptococci), and confounders.



## CHAPTER 2 LITERATURE REVIEW

### Dental Plaque

Dental plaque is an accumulation of various microorganisms adhering to a tooth surface as a biofilm by embedding in the matrix of host pellicles and bacterial origin(19, 20). The mechanisms of dental plaque formation are divided into three steps sequentially: the formation of a pellicle, the occurrence of the pioneer microorganisms and their proliferation, and the collection of filamentous organisms and spirochetes (21). This contributes to host defense(22) while its structure limits the penetration of antimicrobial agents(23, 24). The understanding of dental plaque is crucial because it is also known as the main etiology of dental caries and periodontal disease(1).

Dental plaque genesis begins with the pellicular formation. The pellicle is an acquired acellular material secreting by the host. The major components of the pellicle are glycoproteins(1, 25) combined with other salivary components such as proline-rich proteins, statherin, and fibronectin which facilitate bacterial adhesion(26). The next step is the appearance of the pioneer microorganisms. Gram-positive aerobic bacteria, especially cocci, is the first colonizer, followed by the gram-positive rods and filament after a few days. Then, the bacterial plaque shift toward more anaerobic type like gram-negative organisms and fusiforms(1, 25). Gram-negative bacteria such as *Tannerella forsythia*(27), *Porphyromonas gingivalis*, *Treponema denticola*(19) are pathogenic to dental supportive tissue. When the new microorganisms arrive, they attach to the primary colonizer and implant themselves not only to the preceding salivary components but also to the exopolysaccharides of the bacteria(26). The adhesion between the cell surface and the pellicle is firstly a slight reversible physiochemical interaction which gradually leads to a stronger adhesin-receptor mediated attachment that promotes the secondary colonization(28). The final step is marked by the emergence of spiral forms and spirochetes. However, the final proportion of the species in dental plaque is mostly composed of filamentous organisms because of their abilities to overcome other surrounding cells(20, 21). In the normal conditions, this composition is stable and called

“microbial homeostasis” which endures small stresses such as host defense, oral hygiene, and salivary flow(29). However, whenever the stresses are so critical that the minor components are outgrowing and weaken this stability, the disease such as periodontal disease and dental caries may occur.

In the periodontal aspect, dental plaque induces host response and contributes to the disease. Commonly, if supragingival plaque is not removed for 2-3 weeks, the gingival inflammation will occur(30). During these 3 weeks, the compositions of the dental plaque are more complicated than the beginning. Because dental plaque usually accumulates over the gingival margin and gingival sulcus, the bacteria can utilize nutrients from gingival crevicular fluid together with saliva(31). Moreover, some species even metabolize by-products of another's species. These factors massively exacerbate bacterial growth which is unfavorable to the periodontal tissue. The host response to the bacteria and bacterial products by the inflammatory process resulting in increased vascularity, swelling, leukocytic diapedesis, and the loss of connective tissues(32). Furthermore, once the dental plaque is mineralized and becomes calculus, it harbors more bacteria and strengthens the pathogenicity(25). Finally, if this circumstance is maintained long enough, it possibly results in the periodontal breakdown(33).

Another relevant disease is dental caries. In this point of view, there is a widely accepted concept named “Ecological plaque hypothesis”(34). This concept states that normally there are small amounts of potential pathogens for demineralization such as Mutans streptococci and Lactobacilli even within a healthy host. However, when the environment critically changes, the homeostasis is disturbed. To control the stability of the oral microorganisms, the important factors are host defense, including salivary flow, and composition of the diets(35). Regarding this, the prevention of dental caries highlights on dental plaque removal, together with managing the supplementary risks, for instance, reducing the fermentable sugar, increase salivary flow, and promote remineralization by using fluoride(20).

Orthodontic appliances have an advanced design that promotes plaque accumulation. Complex shapes of bracket wings and slots are likely to be the plaque

retentive area. Evidence showed that patients with fixed orthodontic appliances have higher plaque accumulation(2-6) which harmful to the gingival conditions. Patients with fixed orthodontic appliances were reported for the higher tendency of gingival inflammation and bleeding on probing values(2, 4, 5, 7-9). Some even found more clinical attachment loss in fixed orthodontic patients rather than no treatment group(2). Therefore, the key implication is that it is essential to emphasize plaque control and maintain good oral hygiene, especially in orthodontic patients.

### Oral microorganisms in orthodontic patients

The orthodontic appliance is the critical environmental changes that disturb the oral microbial homeostasis, not only in terms of the higher plaque accumulation but also changing in the contents of the plaque. A study in 1970 by Balenseifen et al (10) showed that after banding and applying archwires for 4-5 weeks the plaque was significantly lower in pH, higher in a concentration of carbohydrates, and higher in the population of cariogenic bacteria which were streptococci and lactobacilli. Mutans streptococci, notified as an important pathogen for dental caries(36), was reported that it increased in the active phase of orthodontic treatment(37). Moreover, the composition of the periodontopathogens in dental plaque are also altered. A longitudinal study by Kim et al in 2012 discovered that there were increases in anaerobic pathogenic bacteria such as *T.forsythia*, *C.rectus*, and *P.nigrescens* in the leveling and alignment phase of fixed orthodontic treatment compared to before the treatment(11). This supported a cross-sectional study of Lee et al in 2005(38). *Aggregatibacter actinomycetemcomitans* was also investigated in several studies and showed the corresponding results(3, 39).

Overall, fixed orthodontic appliances cause higher plaque accumulation, and the microbial shift towards more pathogenic types. It seems that fixed orthodontic patients are vulnerable to the disease like gingival inflammation, periodontitis, and dental caries. With these concerns, orthodontists should motivate their patients to maintain good oral hygiene and evaluate their oral hygiene status regularly.

### Plaque removal in orthodontic patients

There are both mechanical and chemical aspects for domestic plaque removal in orthodontic patients which are toothbrushing, complementary aids (e.g. dental floss, single-tufted brushes), and chemical cleaning.

Toothbrushing is a routine method to remove plaque deposits. It has several techniques based on directions and positions of the bristles such as roll method, horizontal scrub, vibration, and circulation(40). However, the horizontal techniques like scrub or Bass technique were considered to be the preferred method rather than the vertical one(41). A study in fixed orthodontic patients by Nassar et al in 2013 demonstrated that when comparing the efficacy of the reduction of the periodontal indices between brushing with scrub technique, modified Stillman technique, and Bass technique, the Bass technique showed the dramatic reduction of the gingival index rather than other groups(42). This result was sensible because the Bass techniques required placing the bristle on the gingival sulcus in an angular direction, gingival to an archwire, and brushed anteroposteriorly with the short strokes (approximately two teeth). The recommendation by the American Dental Association is that brushing should be performed twice a day with 2 minutes each time(43). Furthermore, the single-tufted brush may be used additionally in case of hard-to-reach areas such as ligature loops or area below springs(44).

Nowadays, power toothbrushes are trying to enhance the efficacy of plaque removal. Although the meta-analysis by Kaklamanos and Kalfas in 2008 showed that there were insufficient comparative studies to claim the effectiveness of power toothbrushes over the manual ones in orthodontic patients(12), a more recent systematic review and meta-analysis in 2017 by Makhmari et al revealed that the power toothbrushes tended to improve gingival condition better than the manual one(13). Furthermore, a study by Erbe et al in 2019 illustrated that the interactive power toothbrushes with Bluetooth technology motivated the adolescents with fixed orthodontic appliances to brush more frequently and reduced more dental plaque(45). However, the

additional comparative controlled trials in orthodontic patients are still necessary to confirm this issue.

#### **Photodynamic therapy, Light-emitting diode (LED) - Blue light and antimicrobial activity**

The inactivation of microorganisms by visible light has been studied for over a century, including in agriculture, industry, and medication. The mechanism of the bactericidal effect of light is based on the photosensitizer, the agent that can absorb the light(46). The bacteria can take the external photosensitizer, besides some strains, even have it endogenously. When the photosensitizer is activated by light with its preferable wavelength, the electrons are transferred to produce the radical ions that react with oxygen and result in cytotoxic species such as superoxide, and hydroxyl radicals (type I reaction)(47). Besides, the type II reaction, the energy is transferred to the ground state triplet oxygen( $^3\text{O}_2$ ) and causes the excited state singlet oxygen( $^1\text{O}_2$ ) which can oxidize bacterial proteins, nucleic acids, and lipids, which are essential for bacterial viability(48). Some periodontal pathogens such as *Prevotella nigrescens* and *Prevotella intermedia* contain endogenous porphyrins(49), one of the photosensitizers. *Streptococcus mutans* had been experimented with many exogenous photosensitizers, for instance, erythrosine, toluidine blue, and malachite green(50-52), even though the specific endogenous photosensitizer has not been clearly revealed.

Light-emitting diode is a light generated by electroluminescence. It is the process that light is released after the current is flown through an inorganic material called semiconductor. In the LED circuit, there are two types of semiconductors; N-type and P-type. The N-type has a large number of electrons, while P-type contains holes. When the current circulates, it raises the energy of the electrons so that they move to combine with the holes, and then the energy is released as photons. The color of the light depends on the material of the semiconductors. For the blue light, gallium nitride(GaN) is mostly used(53).

Blue light is the visible light which its wavelength is about 400-500nm. The effect of blue light on antimicrobial activity was investigated by numerous studies. A study by


Tzung et al in 2004 reported that a 420nm blue light was effective in the treatment of some types of acne. This could be explained by that the *Propionibacterium acne* contained porphyrins that caused the photodynamic effects(54). The comprehensive experiment about the effect of 470nm blue light on 3 bacterial strains and 2 fungi strains(55) indicated that the lethal effects depended on the microbial genus, energy levels, light purity, and temperature. This expanded our understanding that, practically, bacterial strains and environmental factors could affect the outcome. The investigations about the bactericidal effect of blue light on *S. mutans* and oral biofilms suggested that it could reduce the *S. mutans*' viability and inhibit biofilm development(14-17). Anyway, data showed that the blue light could have a lethal effect on the new biofilm formed by *S. mutans* only after 7- or 10-minutes exposure and destroy mostly on the outer surface of the biofilm(14). In 2018, an *in vitro* study by Gomez et al applied 405nm violet-blue light on the *S. mutans* biofilm specimens for 5 minutes twice a day for over 5 days. The result revealed the reduction of bacteria numbers in treated groups(17). Yet, there is no *in vitro* study that uses the blue-light LED toothbrush rather than using the blue light alone. Our question is that if the blue- light LED toothbrush apply on the *S. mutans* biofilm for 2 minutes, like the normal brushing duration, whether it still can inhibit the bacterial growth.

#### Light-emitting diode (LED) toothbrushes

Since light-emitting diode (LED) showed the antimicrobial effects(56, 57), it has been recently applied to toothbrushes in order to enhance the efficacy in reducing dental plaque and improve gingival health. The *in vitro* studies demonstrated that the LED toothbrushes were able to reduce the number of *Porphyromonas gingivalis* attached to titanium and zirconia surfaces(58, 59). In 2015, a pilot clinical study by Genina et al found that the blue LED toothbrushes with 412 nm wavelength significantly reduce dental plaque, gingival bleeding, and inflammation more than the manual toothbrushes(18). Another clinical study by Lee et al 2017 compared the antiplaque and antigingivitis effect between LED electric toothbrushes and the common electric toothbrushes in patients with gingival inflammation or mild periodontitis. The data

showed the significantly lower gingival index in the LED group after 2 weeks and 4 weeks whereas the plaque index did not show any significant differences(60). In this experiment, they used red and white LED lights. Recently, a controlled trial by Kwon et al in 2020 revealed the corresponding results(61). The data presented no significant difference in plaque index at any time points (3 weeks and 6 weeks). However, the gingival index and bleeding on marginal probing were significantly lower in the LED electric toothbrushes group than the non-LED group after 6 weeks. This study used the advanced version called Electric 3-color LED toothbrush which is composed of 2 blue lights, 1 red light, and the other 1 white light.

At present, there is no data about the efficacy of the LED toothbrush in fixed orthodontic patients yet. Our study is willing to fulfill this gap of knowledge which may be useful for orthodontic patients who have difficulties maintaining good oral hygiene.



## CHAPTER 3 MATERIALS AND METHODS

### Participants

The samples of this study were the patients submitted to orthodontic treatment at the Department of Orthodontics, Faculty of Dentistry, Chulalongkorn University. All participants eligible for the study were informed about the objectives and the procedures of the research. Informed written consents had been submitted by all participants before the experimental period. Candidates were recruited by the following criteria.

### Inclusion criteria

1. Patients had been treated with fully bonded fixed appliances, at least from one first molar to another first molar in the same arch, and being treated for more than 1 month with the irregularity index not more than 1 mm(62).
2. Having a minimum of 20 fully erupted permanent teeth excluding the third molars.
3. No known systemic disease, medical condition, or under the medication that affects oral tissues.
4. No supplementary plaque control such as antiseptic mouthwashes or antibiotics for one month before the study.
5. No periodontitis

### Exclusion criteria

1. Patients with systemic disease
2. Taking the medications that influence oral tissues, for example, corticosteroids and nonsteroidal anti-inflammatory drugs.
3. Smoking
4. Apply chemical supplementary plaque control.
5. Patients who have Parkinson's disease or other diseases that might affect hand control.

### Sample size

The sample size was calculated based on gingival index data from a previous study(61) with an alpha of 0.05 and 0.8 power of the test. The result indicated 18



samples. However, to compensate for possible drop out of 30% during the study period, the sample size is adjusted to 24.

#### Allocation technique

Allocation of 24 subject into two groups was done by Block-of-4 randomization.

Group A- Manual/LED

Group B- LED/Manual

There were 6 possibilities to allocate participants to a block equally.

1= AABB, 2= ABAB, 3= BAAB, 4= BABA, 5= BBAA, 6=ABBA

2	6	0	5	8	3	9	9	4	7	6
7	7	0	9	4	4	9	9	5	6	8
1	5	9	5	0	8	2	9	6	3	0
1	7	1	9	9	6	0	8	3	2	9
6	5	4	3	4	4	5	3	5	5	2
2	8	6	2	6	5	7	7	8	8	1
7	4	8	9	0	2	9	0	6	9	8
0	9	3	2	1	6	3	7	5	2	7
0	7	5	0	3	7	8	3	6	6	0
3	6	3	8	0	2	5	2	6	0	6
2	6	4	5	9	9	7	9	4	2	1
3	1	9	0	1	8	5	4	0	2	8
8	3	4	0	5	8	1	9	0	2	3
6	4	3	5	2	7	7	0	8	4	3
7	4	2	7	2	9	9	5	2	4	7
7	5	6	2	7	9	0	9	1	5	2
9	0	9	0	8	0	5	2	4	9	4
2	0	2	7	1	8	7	4	0	8	3
5	2	5	7	2	8	6	0	2	8	5

Figure 2 Table of computer-generated random numbers

Subject	Group	Subject	Group	Subject	Group
1	B	9	A	17	B
2	A	10	A	18	A
3	A	11	B	19	B
4	B	12	B	20	A
5	A	13	B	21	A
6	A	14	B	22	B
7	B	15	A	23	A
8	B	16	A	24	B

Table 1 Group allocation of 25 subjects

A list of random numbers was randomly selected from a table (figure 2), any numbers other than 1 to 6 were excluded. For instance, if the highlighted numbers in figure 2 were selected, the results are 3 1 1 5 4 2. Random allocation was in blocks. (Subject:24, Block size: 4, Group: 2 (A,B))

### **Ethical consideration**

The materials and methods of this study was approved by the Human Research Ethics Committee of the Faculty of Dentistry, Chulalongkorn University, Thailand (approval no.126/2020, study code: HREC-DCU 2020-116), and registered at the Thai Clinical Trials Registry (TCTR; TCTR 20210510004).

### **LED toothbrush**

LED toothbrushes (WHITENGO™, UK) used in our study was certified by European Conformity (CE marking). The specifications include 460-480nm wavelength, 16,000 acoustic pulsations per minute, 840mW of power, and 1,000-3,000 MCD of light intensity. The bristles are made from silicone which were 9mm long. The radiance was  $0.0176 \text{ W/Sr.m}^2$ , measured by spectroradiometer CS-2000 (Konica Minolta, INC, Japan).

### **Research methodology**

#### **1. The clinical part**

This clinical study was conducted as a crossover randomized controlled trial. Screening examinations were performed before starting the experiment. The 24 subjects were randomly allocated into two groups.

Group A: Manual/LED

Group B: LED/Manual

In both groups, the patients were instructed by one dentist for 5 minutes to brush with Bass techniques(42). On the bracket side, patients were also informed to place the bristles to the top and the bottom of the brackets, and brush with horizontal strokes. In group A, the manual toothbrushes were orthodontic toothbrushes by Systema OD (Japan). In group B, the patients brushed with LED toothbrushes (WHITENGO™, UK). The same standard toothpaste was used in both groups. All participants were assigned to brush for a minimum of 2 minutes at least twice a day, in the morning and the evening

after meals. Both groups had been brushing with the assigned toothbrushes for 28 days. After that, the patients will be under the washout interval for the other 28 days. During this period, they will be using their own normal toothbrushes. Then, the patients will switch to use another device for the other 28 days. The accuracy of the brushing method and compliance were confirmed by questionnaires. Other supplementary plaque control methods and scaling were prohibited during these periods.

Diagram of the study design




Figure 3 The cross-over study design flow chart

2. The *in vitro* part (modified from the study of Chebath-Taub et al (14).)

Our study demonstrated the effect of LED toothbrushes on newly formed *S. mutans* biofilms *in vitro* by these following methods.

#### Bacterial strains and growth condition

*Streptococcus mutans* UA 159 from bacterial glycerol stocks were inoculated in Brain-Heart Infusion (BHI) agar, and then were incubated at 37°C with 5% CO<sub>2</sub> for 24 hours. The isolated colony was regrown overnight in BHI broth with sustained shaking at 240 rpm. After that, its optical density at 600nm (OD<sub>600nm</sub>) was evaluated and adjusted until value 0.1. The culture was then incubated at 37°C with 5% CO<sub>2</sub> for 3 hours to reach the determined logarithm phase of growth (OD<sub>600nm</sub> ≈ 0.4-0.6) which would be used for the biofilm formation.

#### Formation of biofilms

The bacterial cells at log phase were harvested by centrifugation (12,000 x g, 4°C, 15 minutes). The cells were then re-suspended in BHI broth with 1% sucrose. The of 3 mL of suspension which contains  $3 \times 10^8$  bacterial cells were added into each plate, and were incubated at 37°C with 5% CO<sub>2</sub> for 36 hours.

#### Exposure of the LED toothbrush to biofilm

After incubation, the supernatant fluid above the biofilm was removed. The biofilm samples were divided into 5 groups depending on the LED exposure time, 2 biofilm plates in each group. In the control group, there was no LED exposure. The other four experimental groups were exposed to the LED toothbrush for 15, 30, 60, and 120 seconds. The LED toothbrush was switched on and held 2 mm above the biofilm.

The remaining attached bacteria were scraped off and put into 100 µl of sterile PBS. The bacteria suspensions were then sonicated and serially diluted ( $10^{-1}$  to  $10^{-8}$ ). The 100 µl of each concentration were dropped onto BHI agar in duplicate, and then were incubated for 36 hours at 37°C with 5% CO<sub>2</sub>. The highest concentration that the colonies can be counted separately in the range of 30-300 colonies were used to calculated the number of bacteria. The percentage of bacterial survival was calculated relative to the control.

The whole experiments were repeated 3 times (total N=6/group)




Figure 4 Flow chart of the in vitro study design

#### Outcome measurement

The plaque index and gingival index were assessed at baseline and after 28 days of each intervention.

#### Examiner calibration

Two trained examiners were calibrated for the evaluation of plaque index and gingival index by assessing both indices on 3 patients with fixed orthodontic appliances. The reliability of the intra-examiner and inter-examiner were evaluated by Kappa statistics. For the plaque index score, the Kappa of 0.861 and 0.887 were obtained for both examiners and a Kappa of 0.874 was obtained from inter examination agreement. For the gingival index score, the Kappa of 0.852 and 0.786 were obtained from both examiners and the Kappa of 0.792 was obtained from inter-examiner agreement.

#### Plaque index (PI)

On the non-bracket sides, Loe-Silness plaque index score(63) was taken. On the bracket sides, the surfaces were divided into 4 zones around the brackets: mesial(M), distal(D), gingival(G), and incisal(I). Each zone was evaluated following Loe-Silness score(64)(Figure 5). The assessment was performed on six teeth according to

Ramfjord's representative teeth(65): maxillary right first molar, maxillary left central incisor, maxillary left first premolar, mandibular left first molar, mandibular right central incisor, and mandibular right first premolar. In case of the first molar is absent or banded, the second molar can substitute. Likewise, the second premolar can replace the missing first premolar.

Score	Criteria
0	No plaque
1	A film of plaque adhering to the free gingival margin and adjacent area of the tooth. The plaque may be seen <i>in situ</i> only after application of disclosing solution or by using the probe on the tooth surface.
2	Moderate accumulation of soft deposits within the gingival pocket, or on the tooth and gingival margin which can be seen with the naked eye.
3	Abundance of soft matter within the gingival pocket and/or on the tooth and gingival margin

Table 2 The Plaque index system

จุฬาลงกรณ์มหาวิทยาลัย  
CHUL




Figure 5 Plaque index bracket.

The modification of Loe-Silness index proposed by Williams. Plaque score of the tooth is divided into mesial (M), distal (D), gingival (G), and incisal (I) parts around the bracket. (Figure from SARUTTICHART, Thayika, et al. Effectiveness of a motionless ultrasonic toothbrush in reducing plaque and gingival inflammation in patients with fixed orthodontic appliances. *The Angle Orthodontist*, 2017, 87.2: 279-285.)(66)

### Gingival index (GI)

Degree of gingival inflammation was rated referring to Loe-Silness gingival index score(67).

Score	Criteria
0	Absence of inflammation
1	Mild inflammation- slight change in color and little change in texture
2	Moderate inflammation- moderate glazing, redness, edema, and hypertrophy Bleeding on pressure
3	Severe inflammation- marked redness and hypertrophy Tendency to spontaneous bleeding Ulceration

*Table 3 Criteria for the Gingival Index Score*

### **Data collection and analysis**

Data was collected and analyzed by using statistical software (IBM SPSS statistics 22, SPSS, Chicago, IL, USA).

In the clinical section, Shapiro-Wilk test was used to quality the normality of the data. Due to the limitation of the study, only 15 patients finished the first period. As a result, we used the available data to analyze as parallel-group. The difference between the value at baseline and after 28 days (within-group) was measured by Wilcoxon Signed Rank test whereas the difference of changes in value between the manual group and the LED group was performed by Mann-Whitney U test.

In the *in vitro* part, Kruskal-Wallis test was applied to test the difference of percentage of bacteria viability among groups.

Statistically significant differences were set at P-value < 0.05.

### Expected benefit of this study

The result of randomized controlled trial with a parallel-group design which assess the effect of LED toothbrush would be beneficial to consider the use of the LED toothbrush for oral care in patients with fixed orthodontic patients.

### Limitation

The result of this study can be applied only in fixed orthodontic patients.

## CHAPTER 4 RESULT

### The clinical part

A total of 22 subjects (5 males and 17 females) with ages ranged from 14 to 54 years (mean  $26.91 \pm \text{SD } 9.86$ ) were enrolled in this study. Not all but 15 participants completed the assigned study period. During the experiment, one female from group B (LED group) had dropped out due to her parents' awareness of LED radiation. Finally, the number of participants included in data analysis was 15, 8 from group A (manual group) and the other 7 from group B (LED group).

### Plaque index score

The presence of visible dental plaque was assessed at baseline and 28 days after brushing with the assigned toothbrush. Table 4 showed the mean PI scores (Pre\_PI and Post\_PI) of the bracket and non-bracket sides. In the manual group, there was no statistical difference between mean PI scores before and after the brushing period in both bracket and non-bracket sides ( $P=0.889$  and  $P=0.228$  respectively). Likewise, in the LED group, there was no statistical difference between mean PI scores before and after the brushing period in both bracket and non-bracket sides ( $P=0.128$  and  $P=0.075$  respectively). When comparing between groups, the changes of the mean PI score were not significantly different in both bracket and non-bracket sides ( $P=0.203$  and  $P=0.412$  respectively). However, when the different tooth surface area of the bracket site was analyzed separately, a significant reduction of the mean PI score was shown in the gingival portion in the LED group ( $P=0.048$ ) whereas there was no significant difference



in the other surface area (Table5). There was also no significant difference in mean PI in the manual group at any surfaces.

Side	Brushing method	N	Pre_PI <sup>a</sup>	Post_PI <sup>b</sup>	<i>P-value</i> <sup>c</sup>	Post-Pre <sup>d</sup>	<i>P-value</i> <sup>e</sup>
Bracket	Manual	8	1.00 (0.22)	0.96 (0.37)	0.889	-0.04 (0.42)	0.203
	LED	7	1.19 (0.33)	0.90 (0.49)	0.128	-0.29 (0.45)	
Non-bracket	Manual	8	1.40 (0.24)	1.23 (0.34)	0.228	-0.08 (0.22)	0.412
	LED	7	1.60 (0.37)	1.19 (0.56)	0.075	-0.19 (0.29)	

*Table 4 Mean (SD) Plaque Index Score (PI) of bracket and non-bracket sides before and after brushing by each method and their differences.*

<sup>a</sup>Pre\_PI = mean PI before brushing by each method

<sup>b</sup>Post\_PI = mean PI after brushing by each method

<sup>c</sup>Wilcoxon Signed Rank test

<sup>d</sup>Post-Pre = mean difference of PI between after and before brushing by each method

<sup>e</sup>Mann-Whitney U test

	Site	Brushing method	N	Pre_PI <sup>a</sup>	Post_PI <sup>b</sup>	<i>P</i> value <sup>c</sup>	Post-Pre <sup>d</sup>	<i>P</i> value <sup>e</sup>
Bracket side	Proximal	Manual	8	1.25 (0.28)	1.35 (0.36)	0.326	0.11 (0.40)	0.164
		LED	7	1.50 (0.41)	1.25 (0.55)	0.248	-0.25 (0.56)	
	Gingival	Manual	8	1.25 (0.22)	1.06 (0.49)	0.256	-0.19 (0.44)	0.180
		LED	7	1.41(0.38)	1.00 (0.52)	0.048*	-0.41 (0.42)	
	Incisal	Manual	8	0.50 (0.30)	0.46 (0.46)	0.799	-0.04 (0.52)	0.486
		LED	7	0.67 0.35)	0.45 (0.49)	0.293	-0.21 (0.45)	

*Table 5 Mean (SD) Plaque Index Score (PI) of a different tooth surface area around the bracket on the bracket side before and after brushing by each method and their differences.*

<sup>a</sup>Pre\_PI = mean PI before brushing by each method

<sup>b</sup>Post\_PI = mean PI after brushing by each method

<sup>c</sup>Wilcoxon Signed Rank test

<sup>d</sup>Post-Pre = mean difference of PI between after and before brushing by each method

<sup>e</sup>Mann-Whitney U test

\*Statistically significant difference (P<0.05)

#### Gingival index score

There was no significant difference detected in the comparisons of gingival index (GI) scores before and after each brushing method or the changes in GI scores between the two groups (Table 6).

Side	Brushing method	N	Pre_GI <sup>a</sup>	Post_GI <sup>b</sup>	<i>P-value</i> <sup>c</sup>	Post-Pre <sup>d</sup>	<i>P-value</i> <sup>e</sup>
Bracket	Manual	8	1.34 (0.20)	1.25 (0.20)	0.285	-0.08 (0.22)	0.553
	LED	7	1.33 (0.27)	1.14 (0.18)	0.121	-0.19 (0.29)	
Non-bracket	Manual	8	1.40 (0.24)	1.29 (0.21)	0.246	-0.11 (0.28)	0.521
	LED	7	1.55 (0.27)	1.31 (0.24)	0.111	-0.24 (0.31)	

Table 6 Mean (SD) Gingival Index Score (GI) of bracket and non-bracket side before and after brushing by each method and their differences.

<sup>a</sup>Pre\_GI = mean GI before brushing by each method

<sup>b</sup>Post\_GI = mean GI after brushing by each method

<sup>c</sup>Wilcoxon Signed Rank test

<sup>d</sup>Post-Pre = mean difference of GI between after and before brushing by each method

<sup>e</sup>Mann-Whitney U test


#### The *in vitro* part

The percentage of bacterial survival was calculated relative to the control. Table 7 showed the percentage of bacterial viability were significantly different among groups ( $P=0.006$ ). After, post hoc comparisons were carried out by Mann-Whitney U test with Bonferroni correction of the significance level of multiple pairwise comparisons ( $p < 0.05$ ), the data indicated that when comparing to the negative control group (no LED), the group with 15, 30, 60 and 120 seconds exposure had significantly lower percentage of bacterial viability than the control group. (Figure 6)

% Bacteria viability		
Intervention	n	Mean (95% CI)
No LED exposure	6	99.98 (62.31, 137.63)
LED 15 seconds	6	25.52 (-1.61, 52.65)
LED 30 seconds	6	17.37 ( -5.43, 40.18)
LED 60 seconds	6	20.02 (-8.77, 48.81)
LED 120 seconds	6	11.28 (1.23, 21.34)
<i>P</i> value <sup>b</sup>		0.006

*Table 7 Percentage of bacterial viability<sup>a</sup>*

<sup>a</sup> % Bacteria viability calculated from the number of bacteria colonies in form of colony-forming unit per milliliter (CFU/mL) when control was adjusted to 100%; CI, confidence interval



*Figure 6 The comparison of percentage of bacterial viability*

## CHAPTER 5 DISCUSSION

This study was analyzed as a parallel-group randomized controlled trial with a 28-day brushing period to compare the effectiveness of the LED toothbrush to the manual toothbrush in terms of reducing dental plaque and gingival inflammation. The randomized design minimizes the biased error for the specific intervention. However, there are concerns regarding the use of parallel-group design because it includes within-subject confounding factors such as age, gender, and personal oral hygiene. We recruited the patients who undergoing fixed orthodontic appliances for more than 1 month because it allowed time for the patients to adapt their brushing skills with the fixed orthodontic appliance, so the outcomes were not influenced by the time-dependent acquisition of brushing skills. Also, we recruited the participants who had less than 1 mm of the remaining crowding. The reason is that the difficulties of brushing by tooth malalignment would be fairly eliminated for all participants. The patients performed their assigned brushing methods at home to illustrate the normal condition. The accuracy of the brushing method and compliance were confirmed by questionnaires.

Our clinical results suggested that neither manual nor LED toothbrushes significantly reduced dental plaque and gingivitis (within-group comparisons). Between-group comparisons showed no significant difference in all indices. However, the LED toothbrush significantly reduced dental plaque at the gingival portion on the bracket side. Although there was no statistical significance, most results indicated the tendency of dental plaque and gingivitis reduction as we can see from the negative mean difference between after and before brushing by each method. Due to the limitation of time and the pandemic of COVID-19, we could not collect the data as we had planned. Comparing to the available data, a study by Genina et al found that the blue LED toothbrushes with 412 nm wavelength significantly reduce dental plaque, gingival bleeding, and inflammation more than the manual toothbrushes after using for one month(18). Resemble our study they used parallel-group randomized study design but

the subjects were 30 persons each group and tested in non-orthodontic patients. Another similar study design by Kwon et al(61) presented that there was no significant difference in plaque index at 3 weeks and 6 weeks while the gingival index and bleeding on marginal probing were significantly lower in the LED electric toothbrushes group than the non-LED group after 6 weeks. This study used the advanced version called Electric 3-color LED toothbrush which is composed of 2 blue lights, 1 red light, and the other 1 white light. Again, this study was done in non-orthodontic patients. Overall, the LED toothbrush is likely to improve gingival health, but there was still no other study in orthodontic patients to conclude these issues, so it is better to collect more data.

The *in vitro* results revealed that the LED light from the LED toothbrush could significantly reduce the *S.mutans* viability in biofilm after 15 seconds exposure. The duration of 120 seconds had the lowest percentage of bacteria viability.

The concept of the bactericidal effect of light is that the bacteria contains photosensitizer, the agent that can absorb the light(46). When the photosensitizer is activated by light with its preferable wavelength, the electrons are transferred to produce the radical ions that react with oxygen and result in cytotoxic species(47). Blue light has been proved to have the bactericidal effect(17). The activity of blue light against *S.mutans* has been shown in previous *in vitro* studies(14, 17, 68). Chebath-Taub et al reported that blue light could reduce *S. mutans* biofilm re-formation(14). Relating to our study, our result supported that the visible blue light from the LED toothbrush could reduce the viability of *S.mutans* in biofilm. However, the light in our study had to pass through the silicon bristles which were 9 mm long. Thus, it is important to test whether the LED light in the toothbrush could still have an antibacterial effect. The results suggested that the light from the LED toothbrushes can significantly reduce *S. mutans* viability in biofilm after a 120-second exposure.

The exposure time is one of the most important factors for the bactericidal effect of the LED toothbrush. Previous *in vitro* studies(14, 17, 68) set the exposure duration from 1 to 10 minutes. However, people usually brush their teeth for 2-3 minutes at a time,

and the toothbrush does not stay at one position for minutes in the patient's mouth. Therefore, our study varied the duration of blue light exposure to find the shortest time that decreased the viability of *S. mutans*. Our data showed that a significant reduction in bacterial viability occurred after the biofilm had been exposed to LED blue light for 15 seconds. However, the LED toothbrush in our study was applied to the biofilm formed in 35x10 mm culture plate, which the size is similar to 2-3 teeth. According to the recommendation of the American Dental Association, toothbrushing should be performed for 2 minutes, twice a day(43). Thus, if an LED toothbrush was applied for 2 minutes, it may help to reduce bacterial viability in the oral cavity.

Overall, our clinical and *in vitro* study added the evidence of the effectiveness of the LED toothbrush. However, further clinical studies with larger participants would be beneficial for more solid evidence especially in fixed orthodontic patients.

#### Conclusion

LED toothbrush was not more effective in reducing dental plaque and gingival inflammation than the manual toothbrush in fixed orthodontic patients. The LED blue light from the LED toothbrush significantly reduced the number of *S.mutans* in biofilm *in vitro* when the biofilm was exposed to the light for at least 15 seconds.

## APPENDICE

### Appendix A

A.1 The individual mean Plaque Index Score (PI) of bracket and non-bracket sides before and after brushing by each method.

Patient No.	Bracket		Non-bracket	
	Pre	Post	Pre	Post
A01	0.89	1.14	1.67	1.50
A02	0.94	0.64	1.50	0.67
A03	1.06	1.39	1.33	1.50
A04	0.78	0.75	1.50	1.33
A05	1.08	0.95	1.17	0.83
A06	1.28	1.50	1.67	1.50
A07	0.67	0.92	1.00	1.50
A08	1.30	0.39	1.33	1.00
B01	1.00	0.95	1.83	1.17
B02	0.94	0.61	1.33	1.00
B03	0.89	0.81	1.67	1.83
B04	1.11	0.72	1.67	1.50
B05	1.33	1.78	1.50	1.67
B06	1.86	1.22	2.17	1.00
B07	1.20	0.22	1.00	0.17



A.2 The individual mean Plaque Index Score (PI) of a different tooth surface area around the bracket on the bracket side before and after brushing by each method.

Patient No.	Proximal		Gingival		Incisal	
	Pre	Post	Pre	Post	Pre	Post
A01	1.17	1.42	1.17	1.00	0.33	1.00
A02	1.33	1.25	1.17	0.67	0.33	0.00
A03	1.17	1.67	1.50	2.00	0.50	0.50
A04	0.83	1.08	1.33	1.17	0.17	0.00
A05	1.25	1.17	1.17	1.00	0.83	0.67
A06	1.67	2.00	1.50	1.33	0.67	1.17
A07	1.00	1.42	0.83	1.00	0.17	0.33
A08	1.58	0.83	1.33	0.33	1.00	0.00
B01	1.17	1.17	1.33	1.00	0.50	0.67
B02	1.33	1.00	1.00	0.67	0.50	0.17
B03	1.00	1.08	1.50	1.17	0.17	0.17
B04	1.67	1.33	1.17	0.83	0.50	0.00
B05	1.67	2.33	1.50	1.83	0.83	1.17
B06	2.25	1.33	2.17	1.33	1.17	1.00
B07	1.42	0.50	1.17	0.17	1.00	0.00

A.3 The individual mean Gingival Index Score (GI) of bracket and non-bracket sides before and after brushing by each method.

Patient No.	Bracket		Non-bracket	
	Pre	Post	Pre	Post
A01	1.50	1.17	1.67	1.50
A02	1.17	1.17	1.50	1.00
A03	1.67	1.67	1.33	1.50
A04	1.17	1.17	1.50	1.50
A05	1.17	1.17	1.17	1.17
A06	1.33	1.33	1.67	1.33
A07	1.17	1.33	1.00	1.33
A08	1.50	1.00	1.33	1.00
B01	1.33	1.00	1.83	1.17
B02	1.17	1.00	1.33	1.17
B03	1.50	1.17	1.67	1.67
B04	1.17	1.00	1.67	1.50
B05	1.33	1.50	1.33	1.50
B06	1.83	1.17	1.83	1.17
B07	1.00	1.17	1.17	1.00

A.4 The geometric mean of the numbers of *S.mutans* and the calculation to percentage of bacterial viability.

Experiment 1	CFU/ml ( $\times 10^9$ )	Mean	%viability	Mean % viability
Control 1	8.48	16.60	51.08	99.97
Control 2	24.71		148.86	
LED 120sec_1	3.17	3.66	19.10	22.05
LED 120sec_2	4.15		25.00	
LED 60sec_1	12.49	7.14	75.24	42.98
LED 60sec_2	1.78		10.72	
LED 30sec_1	0.87	5.32	5.24	32.05
LED 30sec_2	9.77		58.86	
LED 15sec_1	12.85	7.60	77.41	45.75
LED 15sec_2	2.34		14.10	

Experiment 2	CFU/ml ( $\times 10^9$ )	Mean	%viability	Mean % viability
Control 1	8.79	10.96	80.20	99.95
Control 2	13.12		119.71	
LED 120sec_1	0.06	0.15	0.55	1.32
LED 120sec_2	0.23		2.10	
LED 60sec_1	0.53	0.50	4.84	4.52
LED 60sec_2	0.46		4.20	
LED 30sec_1	0.46	0.42	4.20	3.83
LED 30sec_2	0.38		3.47	
LED 15sec_1	2.01	1.57	18.34	14.32
LED 15sec_2	1.13		10.31	

Experiment 3	CFU/ml ( $\times 10^9$ )	Mean	%viability	Mean % viability
Control 1	2.31	1.91	120.94	100.00
Control 2	1.51		79.06	
LED 120sec_1	0.16	0.20	8.38	10.47
LED 120sec_2	0.24		12.57	
LED 60sec_1	0.16	0.24	8.38	12.57
LED 60sec_2	0.32		16.75	
LED 30sec_1	0.46	0.31	24.08	16.23
LED 30sec_2	0.16		8.38	
LED 15sec_1	0.2	0.32	10.47	16.49
LED 15sec_2	0.43		22.51	

## Appendix B

แบบประเมินความร่วมมือระหว่างเข้าร่วมโครงการ

1. เมื่อเข้าร่วมโครงการท่านได้ใช้แปรงสีฟันตามที่ได้รับมอบหมายทุกครั้งหรือไม่

- ☐ ใช้ทุกครั้ง
- ☐ ใช้เป็นส่วนมาก
- ☐ ใช้เป็นบางครั้ง
- ☐ ไม่ได้ใช้เลย

2. ท่านใช้ไหมขัดฟันและ/หรือแปรงซอกฟันหรือไม่

- ☐ ใช้ทุกวัน
- ☐ ใช้เป็นส่วนมาก
- ☐ ใช้เป็นบางครั้ง
- ☐ ไม่เลย

2. ท่านแปรงฟันวันละกี่ครั้ง

- ☐ 1 ครั้ง
- ☐ 2 ครั้ง
- ☐ มากกว่า 2 ครั้ง

3. ท่านใช้ระยะเวลาในการแปรงฟัน

- ☐ ประมาณ 30 วินาที
- ☐ ประมาณ 1 นาที
- ☐ ประมาณ 2 นาที
- ☐ มากกว่า 2 นาที

4. ท่านแปรงฟันด้วยวิธีการตามที่ทันตแพทย์แนะนำหรือไม่ (วางแปรง 45 องศาระหว่างขอบเหงือกและคอฟัน ขยับสั้นๆในแนวราบ 10 ครั้งต่อตำแหน่ง และทำความสะอาดด้านบนและด้านล่างของเครื่องมือจัดฟัน)

- ☐ ทุกครั้ง
- ☐ เป็นส่วนมาก
- ☐ บางครั้ง
- ☐ ไม่เลย



จุฬาลงกรณ์มหาวิทยาลัย  
CHULALONGKORN UNIVERSITY

5. หลังจากแปร่งฟันด้วยแปรงสีฟันที่ได้รับมอบหมาย ท่านรู้สึกว่าการแปรงฟันของท่านมีความสะอาดในระดับใด

- ☐ สะอาดมาก
- ☐ สะอาดปานกลาง
- ☐ สะอาดน้อย

6. โดยภาพรวมท่านมีความพึงพอใจต่อการแปรงสีฟันที่ได้รับมอบหมายในช่วง 1 เดือนที่ผ่านมาในระดับใด

- ☐ มาก
- ☐ ปานกลาง
- ☐ น้อย

7. ท่านใช้น้ำยาบ้วนปากในระหว่างเข้าร่วมโครงการหรือไม่

- ☐ ใช่
- ☐ ไม่ใช่

8. ท่านมีการใช้ยาปฏิชีวนะในระหว่างเข้าร่วมโครงการหรือไม่

- ☐ ใช่
- ☐ ไม่ใช่

9. ท่านได้รับการขูดหินปูนและ/หรือเกลารากฟัน ในระหว่างเข้าร่วมโครงการหรือไม่

- ☐ ใช่
- ☐ ไม่ใช่

## REFERENCES

1. Fine D. Mouthrinses as adjuncts for plaque and gingivitis management. A status report for the American Journal of Dentistry. American journal of dentistry. 1988;1(6):259-63.
2. Zachrisson BU. Cause and prevention of injuries to teeth and supporting structures during orthodontic treatment. American Journal of Orthodontics and Dentofacial Orthopedics. 1976;69(3):285-300.
3. Paolantonio M, Festa F, di Placido G, D'Attilio M, Catamo G, Piccolomini R. Site-specific subgingival colonization by *Actinobacillus actinomycetemcomitans* in orthodontic patients. American Journal of Orthodontics and Dentofacial Orthopedics. 1999;115(4):423-8.
4. Ristic M, Svabic MV, Sasic M, Zelic O. Clinical and microbiological effects of fixed orthodontic appliances on periodontal tissues in adolescents. Orthodontics & craniofacial research. 2007;10(4):187-95.
5. KLOEHN JS, PFEIFER JS. The effect of orthodontic treatment on the periodontium. The Angle Orthodontist. 1974;44(2):127-34.
6. Attack NE, Sandy JR, Addy M. Periodontal and microbiological changes associated with the placement of orthodontic appliances. A review. Journal of Periodontology. 1996;67(2):78-85.
7. Alexander SA. Effects of orthodontic attachments on the gingival health of permanent second molars. American Journal of Orthodontics and Dentofacial Orthopedics. 1991;100(4):337-40.
8. Zachrisson S, Zachrisson BU. Gingival condition associated with orthodontic treatment. Angle orthod. 1972;42(1):26-34.
9. Trossello VK, Gianelly AA. Orthodontic treatment and periodontal status. Journal of periodontology. 1979;50(12):665-71.
10. Balenseifen JW, Madonia J. Study of dental plaque in orthodontic patients. Journal of dental research. 1970;49(2):320-4.

11. Kim S-H, Choi D-S, Jang I, Cha B-K, Jost-Brinkmann P-G, Song J-S. Microbiologic changes in subgingival plaque before and during the early period of orthodontic treatment. *The Angle orthodontist*. 2012;82(2):254-60.
12. Kaklamanos EG, Kalfas S. Meta-analysis on the effectiveness of powered toothbrushes for orthodontic patients. *American Journal of Orthodontics and Dentofacial Orthopedics*. 2008;133(2):187. e1-. e14.
13. Al Makhmari SA, Kaklamanos EG, Athanasiou AE. Short-term and long-term effectiveness of powered toothbrushes in promoting periodontal health during orthodontic treatment: A systematic review and meta-analysis. *American Journal of Orthodontics and Dentofacial Orthopedics*. 2017;152(6):753-66. e7.
14. Chebath-Taub D, Steinberg D, Featherstone JD, Feuerstein O. Influence of blue light on *Streptococcus mutans* re-organization in biofilm. *Journal of Photochemistry and Photobiology B: Biology*. 2012;116:75-8.
15. De Sousa DL, Lima RA, Zanin IC, Klein MI, Janal MN, Duarte S. Effect of twice-daily blue light treatment on matrix-rich biofilm development. *PloS one*. 2015;10(7):e0131941.
16. Gomez GF, Huang R, MacPherson M, Zandona AGF, Gregory RL. Photo inactivation of *Streptococcus mutans* biofilm by violet-blue light. *Current microbiology*. 2016;73(3):426-33.
17. Felix Gomez GG, Lippert F, Ando M, Zandona AF, Eckert GJ, Gregory RL. Effect of violet-blue light on *Streptococcus mutans*-induced enamel demineralization. *Dentistry journal*. 2018;6(2):6.
18. Genina EA, Titorenko VA, Belikov AV, Bashkatov AN, Tuchin VV. Adjunctive dental therapy via tooth plaque reduction and gingivitis treatment by blue light-emitting diodes tooth brushing. *Journal of Biomedical Optics*. 2015;20(12):128004.
19. Socransky SS, Haffajee AD. Dental biofilms: difficult therapeutic targets *Periodontology 2000*. 2002.
20. Marsh PD, editor *Dental plaque as a biofilm and a microbial community—implications for health and disease*. BMC Oral health; 2006: BioMed Central.



21. Bernimoulin JP. Recent concepts in plaque formation. *Journal of clinical Periodontology*. 2003;30:7-9.
22. Marsh PD. Role of the oral microflora in health. *Microbial Ecology in Health and Disease*. 2000;12(3):130-7.
23. Van Steenberghe T, Van Winkelhoff A, De Graaff J. Pathogenic synergy: mixed infections in the oral cavity. *Antonie van Leeuwenhoek*. 1984;50(5-6):789-98.
24. Gilbert P, Maira-Litran T, McBain AJ, Rickard AH, Whyte FW. The physiology and collective recalcitrance of microbial biofilm communities. 2002.
25. Wilkins EM, McCullough PA. *Clinical practice of the dental hygienist*: Lea & Febiger; 1976.
26. Listgarten M. Formation of dental plaque and other oral biofilms. *Dental plaque revisited: oral biofilms in health and disease* Bioline, Cardiff, Wales. 1999:187-210.
27. Sakamoto M, Suzuki M, Umeda M, Ishikawa I, Benno Y. Reclassification of *Bacteroides forsythus* (Tanner et al. 1986) as *Tannerella forsythensis* corrig., gen. nov., comb. nov. *International journal of systematic and evolutionary microbiology*. 2002;52(3):841-9.
28. Kolenbrander PE, Andersen RN. Coaggregation and coadhesion in. *Community structure and co-operation in biofilms*. 2000;59:65.
29. Marsh P, Featherstone A, McKee A, Hallsworth A, Robinson C, Weatherell J, et al. A microbiological study of early caries of approximal surfaces in schoolchildren. *Journal of Dental Research*. 1989;68(7):1151-4.
30. Löe H, Theilade E, Jensen SB. Experimental gingivitis in man. *The Journal of periodontology*. 1965;36(3):177-87.
31. Theilade E, Wright W, Jensen SB, Löe H. Experimental gingivitis in man: II. A longitudinal clinical and bacteriological investigation. *Journal of periodontal research*. 1966;1(1):1-13.
32. Page RC. The role of inflammatory mediators in the pathogenesis of periodontal disease. *Journal of periodontal research*. 1991;26(3):230-42.
33. DePaola LG, Overholser CD, Meiller TF, Minah GE, Niehaus C. Chemotherapeutic

inhibition of supragingival dental plaque and gingivitis development. *Journal of clinical periodontology*. 1989;16(5):311-5.

34. Marsh PD. Are dental diseases examples of ecological catastrophes? *Microbiology*. 2003;149(2):279-94.

35. Marsh P. Host defenses and microbial homeostasis: role of microbial interactions. *J Dent Res*. 1989;68:1567-75.

36. Clarke JK. On the bacterial factor in the aetiology of dental caries. *British journal of experimental pathology*. 1924;5(3):141.

37. Rosenbloom RG, Tinanoff N. Salivary *Streptococcus mutans* levels in patients before, during, and after orthodontic treatment. *American Journal of Orthodontics and Dentofacial Orthopedics*. 1991;100(1):35-7.

38. Lee SM, Yoo SY, Kim H, Kim K, Yoon Y, Lim S, et al. Prevalence of putative periodontopathogens in subgingival dental plaques from gingivitis lesions in Korean orthodontic patients. *JOURNAL OF MICROBIOLOGY-SEOUL*. 2005;43(3):260.

39. Paolantonio M, di Girolamo G, Pedrazzoli V, di Murro C, Picciani C, Catamo G, et al. Occurrence of *Actinobacillus actinomycetemcomitans* in patients wearing orthodontic appliances: A cross-sectional study. *Journal of clinical periodontology*. 1996;23(2):112-8.

40. Zachrisson BU. Oral hygiene for orthodontic patients: current concepts and practical advice. *American Journal of orthodontics*. 1974;66(5):487-97.

41. Sangnes G, Zachrisson B, Gjermo P. Effectiveness of vertical and horizontal brushing techniques in plaque removal. *ASDC journal of dentistry for children*. 1972;39(2):94.

42. Nassar PO, Bombardelli CG, Walker CS, Neves KV, Tonet K, Nishi RN, et al. Periodontal evaluation of different toothbrushing techniques in patients with fixed orthodontic appliances. *Dental press journal of orthodontics*. 2013;18(1):76-80.

43. Association AD. Oral Health Topics AZ. Cleaning Your Teeth and Gums (Oral Hygiene). 2010.

44. James G, Beagrie G. The care of the periodontal tissues during orthodontic treatment. *Dent Pract*. 1963;13:268.

45. Erbe C, Klees V, Braunbeck F, Ferrari-Peron P, Ccahuana-Vasquez RA, Timm H, et al. Comparative assessment of plaque removal and motivation between a manual toothbrush and an interactive power toothbrush in adolescents with fixed orthodontic appliances: A single-center, examiner-blind randomized controlled trial. *American Journal of Orthodontics and Dentofacial Orthopedics*. 2019;155(4):462-72.
46. Soukos NS, Goodson JM. Photodynamic therapy in the control of oral biofilms. *Periodontology 2000*. 2011;55(1):143-66.
47. Athar M, Mukhtar H, Elmetts CA, Zaim MT, Lloyd JR, Bickers DR. In situ evidence for the involvement of superoxide anions in cutaneous porphyrin photosensitization. *Biochemical and biophysical research communications*. 1988;151(3):1054-9.
48. Redmond RW, Gamlin JN. A compilation of singlet oxygen yields from biologically relevant molecules. *Photochemistry and photobiology*. 1999;70(4):391-475.
49. Soukos NS, Som S, Abernethy AD, Ruggiero K, Dunham J, Lee C, et al. Phototargeting oral black-pigmented bacteria. *Antimicrobial agents and chemotherapy*. 2005;49(4):1391-6.
50. Wood S, Metcalf D, Devine D, Robinson C. Erythrosine is a potential photosensitizer for the photodynamic therapy of oral plaque biofilms. *Journal of antimicrobial Chemotherapy*. 2006;57(4):680-4.
51. Zanin IC, Lobo MM, Rodrigues LK, Pimenta LA, Höfling JF, Gonçalves RB. Photosensitization of in vitro biofilms by toluidine blue O combined with a light-emitting diode. *European journal of oral sciences*. 2006;114(1):64-9.
52. Rolim JP, De-Melo MA, Guedes SF, Albuquerque-Filho FB, De Souza JR, Nogueira NA, et al. The antimicrobial activity of photodynamic therapy against *Streptococcus mutans* using different photosensitizers. *Journal of Photochemistry and Photobiology B: Biology*. 2012;106:40-6.
53. Gayral B. LEDs for lighting: Basic physics and prospects for energy savings. *Comptes Rendus Physique*. 2017;18(7-8):453-61.
54. Tzung TY, Wu KH, Huang ML. Blue light phototherapy in the treatment of acne. *Photodermatology, photoimmunology & photomedicine*. 2004;20(5):266-9.

55. De Lucca AJ, Carter-Wientjes C, Williams KA, Bhatnagar D. Blue light (470 nm) effectively inhibits bacterial and fungal growth. *Letters in applied microbiology*. 2012;55(6):460-6.
56. Maclean M, Murdoch LE, MacGregor SJ, Anderson JG. Sporicidal effects of high-intensity 405 nm visible light on endospore-forming bacteria. *Photochemistry and photobiology*. 2013;89(1):120-6.
57. Ghate VS, Ng KS, Zhou W, Yang H, Khoo GH, Yoon W-B, et al. Antibacterial effect of light emitting diodes of visible wavelengths on selected foodborne pathogens at different illumination temperatures. *International journal of food microbiology*. 2013;166(3):399-406.
58. Lee H, Kim Y-G, Um H-S, Chang B-S, Lee SY, Lee J-K. Efficacy of an LED toothbrush on a *Porphyromonas gingivalis* biofilm on a sandblasted and acid-etched titanium surface: an in vitro study. *Journal of Periodontal & Implant Science*. 2018;48(3):164-73.
59. Park JH, Kim Y-G, Um H-S, Lee SY, Lee J-K, Chang B-S. Antimicrobial effect of toothbrush with light emitting diode on dental biofilm attached to zirconia surface: an in vitro study. *Journal of Dental Rehabilitation and Applied Science*. 2019;35(3):160-9.
60. Lee S-J. The effect of Light Emitting Diode electric toothbrush on gingivitis: a randomized controlled trial. *Journal of dental rehabilitation and applied science*. 2017;33(2):119-26.
61. Kwon C, Lee J-M, Suh J-Y, Seo S-J, Lee Y, Kim Y-G. Effects of an electric toothbrush combined with 3-color light-emitting diodes on antiplaque and bleeding control: a randomized controlled study. *Journal of Periodontal & Implant Science*. 2019;50.
62. Little RM. The irregularity index: a quantitative score of mandibular anterior alignment. *American Journal of Orthodontics and Dentofacial Orthopedics*. 1975;68(5):554-63.
63. Silness J, L  e H. Periodontal disease in pregnancy II. Correlation between oral hygiene and periodontal condition. *Acta odontologica scandinavica*. 1964;22(1):121-35.
64. Williams P, Clerehugh V, Worthington H, Shaw W. Comparison of two plaque


indices for use in fixed orthodontic appliance patients. J Dent Res. 1991;70:703.

65. Ramfjord SP. Indices for prevalence and incidence of periodontal disease. 1959.

66. Saruttichart T, Chantarawatit P-o, Leevailoj C, Thanyasrisung P, Pitiphat W, Matangkasombut O. Effectiveness of a motionless ultrasonic toothbrush in reducing plaque and gingival inflammation in patients with fixed orthodontic appliances. The Angle Orthodontist. 2017;87(2):279-85.

67. Loe H, Silness J. Periodontal disease in pregnancy I. Prevalence and severity. Acta odontologica scandinavica. 1963;21(6):533-51.

68. Cohen-Berner J, Steinberg D, Featherstone JD, Feuerstein O. Sustained effects of blue light on *Streptococcus mutans* in regrown biofilm. Lasers in medical science. 2016;31(3):445-52.





จุฬาลงกรณ์มหาวิทยาลัย  
**CHULALONGKORN UNIVERSITY**

## VITA

NAME	Chavirakarn Manphibool
DATE OF BIRTH	25 Nov 1990
PLACE OF BIRTH	Bangkok
INSTITUTIONS ATTENDED	Chulalongkorn University
HOME ADDRESS	9/1017 Ideomix Sukhumvit103 Bld. Bangna district Bangkok Thailand 10260
PUBLICATION	-
AWARD RECEIVED	-

